

Methods for Reducing Metal Artifacts in Computerized Tomography

Yakdiel Rodríguez-Gallo, Rubén Orozco-Morales, Marlen Pérez-Díaz
Universidad Central "Marta Abreu" de Las Villas (UCLV), Cuba
yrodriquez-gallo@uclv.edu.cu

Abstract

Metal artifacts are common in clinical images. Many methods for artifact reduction have been published to overcome this problem. In this work, an image smoothing method for artifact reduction (ISMAR) is proposed for image quality improvement in patients with hip prosthesis and dental fillings, which caused metal artifacts. ISMAR was evaluated and compared with three well-known methods for metal artifact reduction (linear interpolation (LI), normalized metal artifact reduction (NMAR) and frequency split metal artifact reduction (FSMAR)). The new method is based on edge-preserving smoothing via L_0 Gradient Minimization filter. Image quality was evaluated by two experienced radiologists completely blinded to the information about if the image was processed or not to suppress the artifacts. They graded image quality in a five points-scale, where zero is an index of clear artifact presence, and five, a whole artifact suppression. The new method had the best results and it was statistically significant respect to the other tested methods ($p < 0.05$). This new method has a better performance in artifact suppression and tissue feature preservation.

Key words: computerized tomography; image processing; iterative methods; metals; comparative evaluations.

Métodos para reducir los artefactos metálicos en la tomografía computarizada

Resumen

Los artefactos metálicos son comunes en las imágenes clínicas. Muchos métodos para la reducción de los artefactos han sido publicados para superar este problema. En el presente trabajo, un método de suavizado de imágenes para la reducción de artefactos (ISMAR) es propuesto para mejorar la calidad de la imagen en pacientes con prótesis de cadera y empastes dentales, los cuales causaron artefactos metálicos. ISMAR fue evaluado y comparado con otros tres métodos reconocidos por su desempeño en la reducción de los artefactos metálicos (Interpolación lineal (LI), reducción de artefactos de metal normalizados (NMAR) y reducción de artefactos de metal divididos en frecuencia (FSMAR)). El nuevo método se basa en el suavizado y conservación de bordes, utilizando para ello el filtro de minimización de gradiente L_0 . La calidad de la imagen fue evaluada por dos radiólogos experimentados completamente ciegos a la información sobre si la imagen fue procesada o no para suprimir los artefactos. Ellos calificaron la calidad de la imagen en una escala de cinco puntos, donde el cero indica la presencia de artefactos, y el cinco, una supresión total de los artefactos. El nuevo método tuvo los mejores resultados y fue estadísticamente significativo con respecto a los otros métodos probados ($p < 0.05$). Este nuevo método tiene un mejor rendimiento en la supresión de artefactos y en la conservación de las características de los tejidos.

Palabras clave: tomografía computerizada; tratamiento de imágenes; métodos de iteración; metales; evaluaciones comparativas.

Introduction

The presence of metal artifacts has been a major problem in x-ray computed tomography (CT). Metal parts in the field of view attenuate most x-ray photons and generate dark and bright streaks after reconstruction. Metal induced streaking artifacts deteriorate the

diagnostic quality and quantitative value of CT images. They obscure the anatomical structures surrounding the metallic objects and preclude confident diagnosis of the disease [1]–[3].

Different approaches towards metal artifact reduction (MAR) have been published, which can be divided mainly into five classes: Acquisition Improvement, Physics-

based Pre-processing, Projection Completion, Iterative Reconstruction and Image based Approaches [4]. A common approach is to directly correct or replace by interpolation the corrupt projection data in the raw projection's data [5]–[9]. A less study MAR method is to perform post-processing, which aims to correct in the image domain. Post-processing algorithms reduce artifacts after the image has been reconstructed, and do not depend on access to raw projection data [10].

In this paper a novel MAR technique is presented in the image domain that uses image smoothing via L_0 Gradient Minimization method [11], which is assessed in presence of dental fillings and metal hip prosthesis. Finally, the results using a real CT dataset are shown. They demonstrate that the present MAR method can produce better or similar results than LI [12], NMAR [8] and FS-MAR [7], without the use of the raw CT data.

Materials and methods

MAR based on edge-preserving smoothing via L_0 Gradient Minimization filter

The ISMAR method (figure 1) is composed by four steps: metal trace's segmentation (I^{Metal}), prior image generation (I^{Prior}), a projection completion, followed by image reconstruction (I^{Rec}) and finally, a smoothing algorithm is applied to obtain the final result as will be explained subsequently.

A. Metal traces segmentation

In the original image ($I^{Original}$), metals are segmented out based on thresholding. A simple thresholding was applied to find the image metal (I^{Metal}). In this work, 3000 HU is selected as the threshold, which is proper to extract metal according to the literature [13]. Then the forward projection (FP) of the obtained I^{Metal} is performed to get metal traces, which specifies the projections affected by metals.

B. Prior image generation

The main step of the present approach is to generate a good prior image (I^{Prior}), which should contain the most possible soft tissue and bone regions. The MAR methods, which complete sinogram by using forward projection of a prior image, can achieve good results as long as the prior image is close enough to the ground truth. However, most of these MAR methods generate the prior image in a simple way, and the prior image is usually not perfect enough because in some occasions contains also artifacts.

To overcome the above problem, in this work, the I^{Prior} is gotten using a global thresholding fixed between 5HU and 10 HU and a L_0 Gradient Minimization (LOGM) smoothing algorithm, which are applied to the $I^{Original}$. LOGM method was selected due to the mechanism of discretely counting spatial changes, because it can remove low-amplitude structures and globally preserve and enhance outgoing edges, even if they are boundaries of very narrow objects. In this method λ is the




Figure 1. Flow diagram showing the steps of the new MAR based on edge-preserving smoothing via L_0 Gradient Minimization filter algorithm.

smoothing parameter, which controls the degree of smooth, and κ controls the iterations rate. Five iterations are generally performed in the algorithm to get a good I^{prior} , with $\kappa = 1.8$ and $\lambda = 9 \times 10^{-6}$. More iterations could yield an over-smoothing, losing the image information.

C. Projection completion and image reconstruction

In this step, sinogram from I^{Original} , I^{prior} and I^{Metal} were obtained using forward projection. Following the well-known method developed by [8], the original sinogram is normalized (S^{Norm}) by dividing it by the sinogram of the I^{prior} pixelwise. Moreover, all values from the I^{Norm} that lie within the metal trace (I^{Metal}) are replaced in each row by linear interpolation (See Kalender et al. [12] for further details). Afterwards, the corrected sinogram (S^{Corr}) is gotten by denormalization of the interpolated (S^{Inter}). Lastly, the reconstructed image (I^{Rec}) is obtained using filtered back projection (FBP) with linear interpolation and Ram-Lak filter.

D. Smoothing algorithm

I^{Rec} is processed using LOGM smoothing algorithm with two iterations ($\kappa = 1.2$ and $\lambda = 8 \times 10^{-6}$) for bringing out the edges and fine anatomical structures as well as eliminate some remaining streak artifacts due to noise, obtaining the final image (I^{Final}).

Patients

The study was performed by using the DICOM images from CT data sets obtained from 15 patients ranging in age from 21–82 years; and body weight ranged from 39.2–98.9 kg. The eligible criterion was the presence of a metallic implant in the examination area. Only the image data of patients was used in the study, keeping all other patient's information anonymous.

Image acquisition

The CT datasets were obtained for this research from two different CT scanners: a kV on-board imaging (OBI) system integrated in a TrueBeam™ medical linear accelerator (Varian Medical System, Palo Alto, CA) and a Siemens SOMATOM Sensation 16 scanner CT using helical scanning geometry (Santa Clara, Cuba). In both CT scanners the tube voltage was set to 120 kVp. The x-ray tube current was set at 25 mA and 22.6 mA respectively. The matrix of reconstructed image was 512 x 512 pixels, corresponding pixel sizes were 1 mm x 1 mm for OBI and 0.776 mm X 0.776 mm for Siemens SOMATOM Sensation.

Image analysis

Qualitative image analyses

Clinical images obtained by using LI, NMAR, FSMAR and the new algorithm were evaluated by two board certified radiologists independently, each with more than 10 years of clinical experience, blinded to all patient data and image parameters. The images were displayed in random order, and reviewed in the soft tissue window (window level 20 HU, window width 400 HU) and bone window (window level 300 HU, window width 2,500 HU).

The artifacts and the diagnostic interpretability were graded on a five point rating scale by observers. The degree of metal artifacts on images was scored on a scale from 1 to 5 (1, very severe streaks; 2, severe streaks; 3, moderate streaks; 4, minimal streaks; 5, no streak artifacts). The diagnostic image quality was similarly scored on CT images on a scale from 1 to 5 (1, severely reduced image quality, non-diagnostic; 2, markedly reduced image quality, with impaired diagnostic interpretability; 3, acceptable image quality and diagnostic interpretability; 4, good image quality, with high diagnostic confidence; 5, excellent image quality, with full diagnostic interpretability).

Statistical analysis

All numeric values were reported as the mean $\pm$ SD. To compare image quality scores of the data sets, non-parametric Friedman, Wilcoxon signed-rank (paired samples) and Mann-Whitney (unpaired samples) tests were employed. To assess inter-observer agreement was used Cohen's kappa. The κ values of 0.01–0.20 were considered to indicate slight agreement, 0.21–0.40 fair agreement, 0.41–0.60 moderate agreement, 0.61–0.80 substantial agreement and 0.81–1.00 almost perfect agreement.

Statistical analyses were performed with statistical software (SPSS, version 22.0; IBM, Chicago, IL, USA). For all statistical analyses, p-values less than 0.05 were considered to represent statistically significant differences.

Results

Artifacts from metallic devices were significantly lower in all algorithms compared with the original image (figure 2), obtained using FBP from raw data. ISMAR led a significant artifact reduction and tissue preservation compared with LI and NMAR. FSMAR and ISMAR had similar results in some cases, but in presence of severe artifacts in the dental fillings, ISMAR had better results. Two examples of images from our study are shown in figure 3.




Figure 2. Comparison of metal artifact reduction and images quality scores between MAR algorithms obtained from non-parametric Friedman test, existing significant difference among methods both hip prosthesis as well as dental fillings with $p < 0.001$. IQ an AR represent image quality and artifacts reduction respectively.




Figure 3. Patient 1 (P1) with two big dental fillings distributed in the right side of the lower jaw and patient 2 (P2) with bilateral hip prosthesis (a) The original artifact CT image (FBP) and the corrected images using the various MAR approaches: (b) LI, (c) NMAR, (d) FSMAR and (e) ISMAR method. The window width and window level was (WW =2500 HU, WL=300 HU).

From table 1 the results on image quality obtained, shown that in presence of dental implants there was significant difference among methods. The Wilcoxon test did not show statistically significant among NMAR and FSMAR ($p=0.271$), as well as among FSMAR and ISMAR ($p=0.058$).

Table 1. Differences in image quality among the algorithms obtained from subjective analysis using Wilcoxon signed-rank (paired samples) and Mann–Whitney (unpaired samples) tests.

Dental implants				
	LI	NMAR	FSMAR	ISMAR
FBP	< 0.001	< 0.001	< 0.001	< 0.001
LI		0.003	0.002	< 0.001
NMAR			0.005	< 0.001
FSMAR				0.009
Hip implants				
	LI	NMAR	FSMAR	ISMAR
FBP	< 0.001	0.001	< 0.001	< 0.001
LI		0.008	0.002	0.001
NMAR			0.271	0.015
FSMAR				0.058

Interobserver agreement for all assessments of image quality is shown in table 2. The κ values for the two observers ranged from 0.681 to 0.779 for image quality of the dental fillings ($P < 0.01$) and from 0.703 to 0.831 for image quality of the hip prosthesis ($P < 0.01$); indicating that substantial to almost perfect interobserver agreements were obtained.

Discussion

Artifacts related to metal hardware are a frequent problem in CT. In this paper, a new method for the suppression of metal artifacts was proposed. The main goal of ISMAR was the creation of a prior image using a LOGM smoothing algorithm, because a good prior image can avoid wrong tissue classification and consequently a better final image.

On the other side, the major metal streak artifacts are corrected in the output image of the proposed algorithm. A prior image ensures that traces of high-contrast structures are continued through the metal trace. Therefore, the typical artifacts tangent to the metal implants,

Table 2. Interobserver agreement for all image quality assessments and all visualization capability assessments.

	Original Image		LI		NMAR		FSMAR		ISMAR	
	κ	P	κ	P	κ	P	κ	P	κ	P
Hip prosthesis	0.831	< 0.01	0.737	< 0.01	0.703	< 0.01	0.722	< 0.01	0.818	< 0.01
Dental fillings	0.758	< 0.01	0.681	< 0.01	0.762	< 0.01	0.779	< 0.01	0.748	< 0.01

and connecting metal implants and high-contrast structures, are reduced to a minimum or even removed completely.

The combination of linear interpolation and the normalization approach led to a quick and efficient reconstruction performance. ISMAR also has advantages for clinical work flow. It has the ability to operate directly on DICOM images, reducing the artifacts and improving image quality. Many MAR algorithms require access to raw projection data, but these often cannot be implemented on commercial CT scanners due to restrictions. From this algorithm could develop a dedicated software for use in workstations by radiologists, even for personal laptops.

The main limitation of this contribution is the number of patients studied. It would be important increase the sample for evaluating its performance, as well as investigate other common sources of streaking artifacts, such as seed implants, spinal screws and surgical clips.

Conclusion

In this work, we presented a novel method for metal artifact reduction in CT images based on edge-preserving smoothing via L0 Gradient Minimization filter. In clinical case studies, the algorithm developed is able to improve image quality. Results obtained using real datasets show that ISMAR is effective reducing metal artifacts and also producing better segmentation, without the use of the raw CT data.

Acknowledgment

The authors would like to thank Professor Dr. Norbert Pelc from Stanford University and Dr. Sam Mazin from RefleXion Medical for proving a clinical data.

References

- [1] ABDOLI M, MEHRANIAN A, AILIANOU A, et. al. Comparative methods for metal artifact reduction in x-ray CT. IEEE Nuclear Science Symposium and Medical Imaging Conference (NSS/MIC). 8-15 Nov. 2014. Seattle, WA, USA. p. 1-5.
- [2] WELLENBERG RHH, et. al. Quantifying metal artefact reduction using virtual monochromatic dual-layer detector spectral CT imaging in unilateral and bilateral total hip prostheses. Eur J Radiology. 2017; 88: 61-70.
- [3] GIANTSOURDI D, et. al. Metal artifacts in computed tomography for radiation therapy planning: dosimetric effects and impact of metal artifact reduction. Phys Med Biol. 2017; 62(8): R49-R80.
- [4] GJESTEBY L, et. al. Metal artifact reduction in CT: where are we after four decades?. IEEE Access. 2016; 4: 5826-5849.
- [5] MOUTON A, MEGHERBI N, VAN SLAMBROUCK K, et. al. An experimental survey of metal artefact reduction in computed tomography. J Xray Sci Technol. 2013; 21(2): 193-226.
- [6] ZHANG H, WANG L, LI L, et. al. Iterative metal artifact reduction for x-ray computed tomography using unmatched projector/backprojector pairs. Medical Physics. 2016; 43(6): 3019-3033.
- [7] MEYER E, RAUPACH R, LELL M, et. al. Frequency split metal artifact reduction (FSMAR) in computed tomography. Medical physics. 2012; 39(4): 1904-1916.
- [8] MEYER E, RAUPACH R, LELL M, et. al. Normalized metal artifact reduction (NMAR) in computed tomography. Med. Phys. 2010; 37(10): 5482-5493.
- [9] GLOVER GH, PELC N J. An algorithm for the reduction of metal clip artifacts in CT reconstructions. Medical physics. 1981; 8(6): 799-807.
- [10] GOLDEN C, et. al. A comparison of four algorithms for metal artifact reduction in CT imaging. 2011; 7961: 79612Y-79612Y-12.
- [11] XU L, LU C, XU Y, JIA J. Image Smoothing via L0 Gradient Minimization. ACM Trans. Graph. 2011; 30(6): 174:1-174:12.
- [12] KALENDER WA, HEBEL R, EBERSBERGER J. Reduction of CT artifacts caused by metallic implants. Radiology. 1987; 164(2): 576-577.
- [13] ZHANG Y, YAN H, JIA X, et. al. A hybrid metal artifact reduction algorithm for x-ray CT. Med. Phys. 2013; 40: 4.
- [14] XU L, LU C, XU Y, JIA J. Image smoothing via L0 gradient minimization. Proceedings of the 2011 SIGGRAPH Asia Conference on - SA '11, vol. 30, no. 6, p. 1, 2011.

Recibido: 13 de febrero de 2018

Aceptado: 29 de mayo de 2018